



America

# CERTIFICATE

No. U8 18 05 69850 002

**Holder of Certificate:** Heidolph Instruments GmbH & Co. KGWalpersdorfer Str. 12  
91126 Schwabach  
GERMANY**Production Facility(ies):** 69850**Certification Mark:****C US****Product:** Laboratory Equipment**Model(s):** MR Hei-Standard  
MR Hei-Tec  
MR Hei-Connect**Parameters:**  
Rated Voltage: 115 or 230 VAC  
Rated Frequency: 50-60 Hz  
Rated Power: 625 W for 115 V devices  
825 W for 230 V devices  
Protection Class: I**Tested according to:**  
UL 61010-1:2012/R:2016-04  
CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04  
UL 61010-2-010:2015  
CAN/CSA-C22.2 NO. 61010-2-010:2015  
UL 61010-2-051:2015  
CAN/CSA-C22.2 No. 61010-2-051:2015

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

**Test report no.:** 028-713125243-000**Date,** 2018-05-25

Page 1 of 1

